



F. BOSCH  
M E D I Z I N T E C H N I K

Friedrich Bosch GmbH & Co. KG, · Postfach 5 · D-72402 Bisingen

Hohenlaienstraße 30  
D-72406 Bisingen  
Telefon 0 74 76 / 35 54, 35 55  
Telefax 0 74 76 / 35 56  
e-mail: jb@fbosch.de  
Internet: http://www.FBosch.de

## DECLARATION OF CONFORMITY

We company **FRIEDRICH BOSCH GMBH&CO.KG, Hohenlaienstr. 30, 72406 Bisingen, Germany**

declare under our sole responsibility that the product:

aneroid sphygmomanometer Prakticus II

articlenumber: 9022.20.112	Prakticus II with velcro cuff black
9022.20.116	Prakticus II with velcro cuff red
9022.20.117	Prakticus II with velcro cuff blue
9022.20.108	Prakticus II with velcro cuff green
9022.20.109	Prakticus II with velcro cuff yellow
9022.20.110	Prakticus II with velcro cuff orange

class of product: class **I M**

to which this declaration relates is in conformity with the following standards  
**Guideline 93/42 EWG appendix VI.3.**

The quality system is attested by the accredited body **TÜV PRODUCT SERVICE** with the licence number **0123**.

The mentioned product is manufactured in compliance with EN 1060-1 and 2: 1994.

The quality system meets the requirements of EN 46003.

date: 30.01.06

Josef Bieger  
Director of quality-management



Registergericht Hechingen HRA 693  
Verwaltungsgesellschaft mbH,  
Sitz Bisingen HRB 287  
Geschäftsführer: Josef Bieger  
Dipl.-Kaufmann