



REBOTEC Rehabilitationsmittel GmbH
D-49610 Quakenbrück, Artlandstr. 57-59
Telefon: 05431/9416-0
Telefax : 05431/9416-66
<http://www.rebotec.de>
e-mail: info@rebotec.de



DECLARATION OF CONFORMITY

We, REBOTEC Rehabilitation GmbH, hereby declare that the following product(s):

- | | |
|--------------------|--|
| Art. no. 180.00 | “POLO-Standard” aluminium rollator |
| Art. no. 181.00 | “POLO-PLUS” aluminium rollator |
| Art. no. 182... | “POLO-Classic” steel rollator |
| Art. no. 188.00 | “JUMBO“ steel rollator up to 150 kg body weight |
| Art. no. 184... | “LUPO” rollator with patented step assistance |
| Art. no. 186.00.10 | “FIX“ aluminium rollator for children |
| Art. no. 186.00.20 | “FOXI” aluminium rollator for young people |
| | |
| Art. no. 187.00.10 | “Pluto-Mini“ children reverse walker different sizes |
| Art. no. 187.00.20 | “Pluto-Maxi“ children reverse walker different sizes |

that is covered by this declaration is in accordance with
the following standard(s) or other normative document(s)

**EN 12182 Technical aids for disabled persons –
General requirements and test methods**, where the following standards are used:
**EN ISO 14971 Medical devices –
Application of risk management to medical devices**

This declaration corresponds with the demands in MDD 93/42-EEC

All products are designed for disabled person's use in hospitals, nursing and private homes.

Purpose

The Rollator (mobile walking-aid) is usable for the interior and external area (indoors/outdoors – traffic). It's increasing the mobility of people with reduced walking ability. The reverse walker is suitable for a supervised indoor walking training for children with reduced walking ability.

Date: 02th January, 2007

i.A. U. Westphal

safety officer

The declaration is not valid any more if the products mentioned above get changed without the permission of REBOTEC.

This declaration of conformity was drawn up automatically and is valid without signature.